

Pathlab Bay of Plenty / Waikato / Whakatane & Rotorua – Reference Intervals 2017

	Lab Test Name	Lab Test Abbreviation	Gender <i>(Male, Female or Both)</i>	Age Low <i>(Years)</i>	Age High <i>(Years)</i>	Low Range	High Range	Lab Units
Coagulation tests	International Normalised Ratio	INR	Both	14	99	0.8	1.2	INR
	Activated partial thromboplastin time	APTT	Both	14	99	24.5	35.5	seconds
	Thrombin clotting time	TCT	Both	14	99	20.0	27.0	seconds

Alan Neal

Signed:

Name: Alan Neal

Position: Lead of Speciality Haematology

Date: 17th February 2017